	Application No.	Applicant(s)	
Notice of Allowability	10/619,690	KIENZLE ET AL.	
	Examiner	Art Unit	
	Johnnie L Smith II	2881	
The MAILING DATE of this communication appears on the cover sheet with the correspondence address All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.			
1. This communication is responsive to <u>amendment filed 09/30/2004</u> .			
2. The allowed claim(s) is/are <u>1-40</u> .			
3. The drawings filed on are accepted by the Examiner.			
 4. Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some* c) None of the: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. 3. Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)). * Certified copies not received: 			
Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application. THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.			
5. A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.			
 6. CORRECTED DRAWINGS (as "replacement sheets") must be submitted. (a) including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached 1) hereto or 2) to Paper No./Mail Date (b) including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d). 			
7. DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.			
Attachment(s) 1. Notice of References Cited (PTO-892) 2. Notice of Draftperson's Patent Drawing Review (PTO-948) 3. Information Disclosure Statements (PTO-1449 or PTO/SB/0 Paper No./Mail Date	5. ☐ Notice of Informal P 6. ☐ Interview Summary Paper No./Mail Dat 7. ☐ Examiner's Amendn 8. ☑ Examiner's Stateme 9. ☐ Other	(PTO-413), enent/Comment	. ·

Application/Control Number: 10/619,690 Page 2

Art Unit: 2881

DETAILED ACTION

Allowable Subject Matter

- 1. Claims 1-40 are allowed.
- 2. The following is an examiner's statement of reasons for allowance: as stated in the Detailed Office Action mailed 06/30/2004, in reference to claims 1-34, the prior art searched and cited failed to clearly teach or fairly suggest, a particleoptical arrangement, having a beam changing arrangement comprising a comb lens having at least one row of a plurality of field source members for providing a common spatial region with at least one of an electric field and a magnetic field, for acting on the beams entering the common field region, and wherein the row extending in the first direction as disclosed in claims 1 and 29. The prior art searched and cited failed to teach or fairly suggest a particle-optical deflecting arrangement for deflecting at least one beam of charged particles, wherein the deflecting arrangement comprises: a comb lens having at least one row of a plurality of field source members, the row extending in the first direction, and a driver arrangement for controlling source strengths of the individual field source members such that at least a dipole field is generated in a field region passed by the beam, the dipole field deflecting the beam as disclosed in claims 15, 19 and 26.

With respect to claim 35, the prior art searched and cited failed to teach of fairly suggest a lithographic method having steps of transferring a pattern onto a particlesensitive substrate, wherein secondary electrons are released from the particlesensitive substrate by the at least one writing beam and emanate from the particlesensitive substrate, and detecting the secondary electrons with at least one electronic detector. Claim 46 is allowable because of its dependencies upon claim 35. Newly added claims 37-40 do not introduce elements and limitations that are clearly found in the prior art searched, accordingly, the claims are considered allowable.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Johnnie L Smith II whose telephone number is 571-272-2481. The examiner can normally be reached on Monday-Thursday 7-4 P.M. and Alternate Fridays.

Page 4

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, John R Lee can be reached on 571-272-2477. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Johnnie L Smith II
Examiner

Art Unit 2881

II SII

SUPERVISORY PATERIT FXAMINE